

Becker Orthopedic

WRIST JOINTS

These small and lightweight joints (distal bar -2", proximal bar-4") are useful for upper extremity applications. Adjustable ROM joint is attached with Chicago screws/posts through slots. Plastic joint can be attached with Velcro® (enclosed), or holes can be drilled.

If used to manage spasticity or contractures, two aluminum ROM joints are recommended instead of one aluminum and one plastic joint. When the plastic joint is used in applications requiring increased strength or control, attach the bar with rivets or screws.

After setting the correct ROM, remove orthosis before tightening the screws with an allen wrench. Loctite® screws in place if ROM will stay constant.

Item No.	Description
U-45	Range of Motion Wrist Joint



(800) 521-2192 (248) 588-7480

BeckerOrthopedic.com

Rev A 6-9-09

Becker Orthopedic

WRIST JOINTS

These small and lightweight joints (distal bar -2", proximal bar-4") are useful for upper extremity applications. Adjustable ROM joint is attached with Chicago screws/posts through slots. Plastic joint can be attached with Velcro® (enclosed), or holes can be drilled.

If used to manage spasticity or contractures, two aluminum ROM joints are recommended instead of one aluminum and one plastic joint. When the plastic joint is used in applications requiring increased strength or control, attach the bar with rivets or screws.

After setting the correct ROM, remove orthosis before tightening the screws with an allen wrench. Loctite® screws in place if ROM will stay constant.

Item No.	Description
U-45	Range of Motion Wrist Joint



(800) 521-2192 (248) 588-7480

BeckerOrthopedic.com

Rev A 6-9-09

Becker Orthopedic

WRIST JOINTS

These small and lightweight joints (distal bar -2", proximal bar-4") are useful for upper extremity applications. Adjustable ROM joint is attached with Chicago screws/posts through slots. Plastic joint can be attached with Velcro® (enclosed), or holes can be drilled.

If used to manage spasticity or contractures, two aluminum ROM joints are recommended instead of one aluminum and one plastic joint. When the plastic joint is used in applications requiring increased strength or control, attach the bar with rivets or screws.

After setting the correct ROM, remove orthosis before tightening the screws with an allen wrench. Loctite® screws in place if ROM will stay constant.

Item No.	Description
U-45	Range of Motion Wrist Joint



(800) 521-2192 (248) 588-7480

BeckerOrthopedic.com

Rev A 6-9-09

APPLICATION/FITTING INSTRUCTIONS

The application instructions/fitting guidelines for this product are for the use of and by certified, qualified orthopedic professionals only. Patients are not to attempt to apply or adjust any brace unless expressly instructed to do so by the orthopedic professional responsible for the prescription and or initial fitting of the orthosis. All questions should be referred to the orthopedic professional and not to the manufacturer. The manufacturer warrants only that the enclosed orthosis has been inspected for quality and can be effective for certain indications, but final decisions and ongoing monitoring must be made by the orthopedic professional(s) prescribing and/or fitting the orthosis to determine its effectiveness for an individual patient. Patient compliance is an integral part of the entire orthotic protocol and must be adhered to in order to avoid potential problems and to maximize the effectiveness of the prescribed treatment.

FOR SINGLE PATIENT USE ONLY

As a condition of the sale of this product, the orthosis is restricted to a “Single Use Only” by the originally fitted patient in order to protect the care provider and the patient against potentially adverse consequences of infectious disease transmission, material instability in adapting to the configuration of the original user and/or decrease in orthotic therapy. Any express or implied warranties are voided if the orthosis is attempted to be reused. Additionally, a license of right to use under any relevant patents pertaining to the orthosis is terminated with the cessation of use by the original patient. As with all our products, this orthosis must be prescribed and applied by a qualified orthopedic professional to determine it meets the needs of the particular patient and accomplishes the desired results.

APPLICATION/FITTING INSTRUCTIONS

The application instructions/fitting guidelines for this product are for the use of and by certified, qualified orthopedic professionals only. Patients are not to attempt to apply or adjust any brace unless expressly instructed to do so by the orthopedic professional responsible for the prescription and or initial fitting of the orthosis. All questions should be referred to the orthopedic professional and not to the manufacturer. The manufacturer warrants only that the enclosed orthosis has been inspected for quality and can be effective for certain indications, but final decisions and ongoing monitoring must be made by the orthopedic professional(s) prescribing and/or fitting the orthosis to determine its effectiveness for an individual patient. Patient compliance is an integral part of the entire orthotic protocol and must be adhered to in order to avoid potential problems and to maximize the effectiveness of the prescribed treatment.

FOR SINGLE PATIENT USE ONLY

As a condition of the sale of this product, the orthosis is restricted to a “Single Use Only” by the originally fitted patient in order to protect the care provider and the patient against potentially adverse consequences of infectious disease transmission, material instability in adapting to the configuration of the original user and/or decrease in orthotic therapy. Any express or implied warranties are voided if the orthosis is attempted to be reused. Additionally, a license of right to use under any relevant patents pertaining to the orthosis is terminated with the cessation of use by the original patient. As with all our products, this orthosis must be prescribed and applied by a qualified orthopedic professional to determine it meets the needs of the particular patient and accomplishes the desired results.

APPLICATION/FITTING INSTRUCTIONS

The application instructions/fitting guidelines for this product are for the use of and by certified, qualified orthopedic professionals only. Patients are not to attempt to apply or adjust any brace unless expressly instructed to do so by the orthopedic professional responsible for the prescription and or initial fitting of the orthosis. All questions should be referred to the orthopedic professional and not to the manufacturer. The manufacturer warrants only that the enclosed orthosis has been inspected for quality and can be effective for certain indications, but final decisions and ongoing monitoring must be made by the orthopedic professional(s) prescribing and/or fitting the orthosis to determine its effectiveness for an individual patient. Patient compliance is an integral part of the entire orthotic protocol and must be adhered to in order to avoid potential problems and to maximize the effectiveness of the prescribed treatment.

FOR SINGLE PATIENT USE ONLY

As a condition of the sale of this product, the orthosis is restricted to a “Single Use Only” by the originally fitted patient in order to protect the care provider and the patient against potentially adverse consequences of infectious disease transmission, material instability in adapting to the configuration of the original user and/or decrease in orthotic therapy. Any express or implied warranties are voided if the orthosis is attempted to be reused. Additionally, a license of right to use under any relevant patents pertaining to the orthosis is terminated with the cessation of use by the original patient. As with all our products, this orthosis must be prescribed and applied by a qualified orthopedic professional to determine it meets the needs of the particular patient and accomplishes the desired results.